UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/516,434	11/30/2004	Joseph P. Orban III	2846	2941
Covidien	7590 05/12/200	9	EXAMINER	
60 Middletown North Haven, (BUI, VY Q	
North Haven, C	.1 00473		ART UNIT	PAPER NUMBER
			3773	
			MAIL DATE	DELIVERY MODE
			05/12/2009	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

		Application No.	Applicant(s)				
Office Action Summary		10/516,434	ORBAN, JOSEPH P.				
		Examiner	Art Unit				
		Vy Q. Bui	3773				
Period fo	The MAILING DATE of this communication ap or Reply	pears on the cover sheet with the c	correspondence address				
WHIC - Exter after - If NC - Failu Any (ORTENED STATUTORY PERIOD FOR REPLEHEVER IS LONGER, FROM THE MAILING Desions of time may be available under the provisions of 37 CFR 1.1 SIX (6) MONTHS from the mailing date of this communication. It is period for reply is specified above, the maximum statutory period reto reply within the set or extended period for reply will, by statute reply received by the Office later than three months after the mailing and patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 136(a). In no event, however, may a reply be tin will apply and will expire SIX (6) MONTHS from e, cause the application to become ABANDONE	N. nely filed the mailing date of this communication. D (35 U.S.C. § 133).				
Status							
1) 又	Responsive to communication(s) filed on 23 F	ehruary 2009					
•		s action is non-final.					
3)	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is						
٥,١	closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.						
Dispositi	on of Claims	,					
· · ·	Claim(s) <u>1-3, 5-12</u> is/are pending in the applic	ation					
•	4a) Of the above claim(s) is/are withdrawn from consideration.						
	5) Claim(s) is/are allowed.						
· ·	S) Claim(s) 1-3, 5-12 is/are rejected.						
•	Claim(s) is/are objected to. Claim(s) are subject to restriction and/o	or election requirement					
ا (۵	claim(s) are subject to restriction and/c	or election requirement.					
Applicati	on Papers						
9)☐ The specification is objected to by the Examiner.							
10)	The drawing(s) filed on is/are: a)∏ acc	cepted or b) \square objected to by the I	Examiner.				
	Applicant may not request that any objection to the	drawing(s) be held in abeyance. See	e 37 CFR 1.85(a).				
	Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).						
11)	11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
Priority ι	ınder 35 U.S.C. § 119						
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 							
2) Notic 3) Inform	e of References Cited (PTO-892) e of Draftsperson's Patent Drawing Review (PTO-948) nation Disclosure Statement(s) (PTO/SB/08)	4) Interview Summary Paper No(s)/Mail Da 5) Notice of Informal P	ate				
Paper No(s)/Mail Date 6) Uther:							

DETAILED ACTION

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claim 1 recites the limitation "the locking leg portion" in lines 9-11. There is insufficient antecedent basis for this limitation in the claim.

Claim Rejections - 35 USC § 102

The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

Claims 1-3, 5-7, 9-14 are rejected under 35 U.S.C. 102(b) as being anticipated by Moufarrege- 5,591,206.

As to claims 1-2, and 6-10, Moufarrege- 5,591,206 (Fig. 1-7; col. 7, lines 8-11) discloses a approximating device comprising 1st fastener male portion 24 of a poly-glycolic acid polymer (having locking leg portion 44 with projections between recesses 58 as a fixing element and anchoring leg portion 56), 2nd fastener female portion 22 of a poly-glycolic acid polymer (having locking leg portion 26 with projections 40 as a fixing element and anchoring leg portion 38), 1st member 74 engaging 1st fastener male portion 24 and 2nd member 74 engaging 2nd fastener female portion 22 (Fig. 4-5) substantially as recited n the claims.

As to claims 3 and 13, anchoring legs 56 and 38 of 1st and 2nd fastener portions includes sharp tips (Fgi. 2).

As to claim 11, Moufarrege- 5,591,206 (Fig. 6-7; col. 5, lines 54-56) discloses 1st member 74 of a hook shape (parabolic shape) served as an anvil to engage lip portion (the portion where locking leg portion 44 meets anchoring leg portion 56 as shown in Fig. 2) of 1st male portion.

As to claim 12, Moufarrege- 5,591,206 (Fig. 6-7; col. 5, lines 54-56) discloses 2nd member 74 of a cylindrical shape having a recess/lumen to engage lip portion (the portion where locking leg portion 38 meets anchoring leg portion 26) of 2nd female portion (Fig. 1-2).

Claim Rejections - 35 USC § 103

The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

Claim 8 is rejected under 35 U.S.C. 103(a) as being unpatentable over Moufarrege-5,591,206 in view of Hogendijk et al-6,051,007.

As to claim 8, Moufarrege- 5,591,206 discloses a approximating device substantially as recited in the claim, except for two pairs of walls forming a rectangular locking passage.

However, Hogendijk-'007 (Fig. 1) discloses rectangular locking passage 25. In view of Hogendijk-'007, it would have been obvious to one of ordinary skill in the art to make a Moufarrege- 5,591,206 approximating device having a rectangular locking passage as recited in the claim as this configuration is also another suitable design for Moufarrege- 5,591,206 device.

Art Unit: 3773

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jackie Ho can be reached on 571-272-4696. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Vy Q. Bui/

Primary Examiner, Art Unit 3773